

Pierce[®] Protein A/G Magnetic Beads

88802 88803

2339.1

Number	Description
88802	Pierce Protein A/G Magnetic Beads, 1mL, supplied at 10mg/mL in water containing 0.05% NaN ₃
88803	Pierce Protein A/G Magnetic Beads, 5mL, supplied at 10mg/mL in water containing 0.05% NaN ₃

Storage: Upon receipt store at 4°C. Product shipped with an ice pack.

Table of Contents

Introduction	1
Important Product Information	2
Procedure for Manual Antibody Purification.....	2
Procedure for Automated Antibody Purification	3
Procedure for Manual Immunoprecipitation.....	4
Procedure for Automated Immunoprecipitation	5
Troubleshooting.....	6
Additional Information	6
Frequently Asked Questions for the Thermo Scientific KingFisher Instrument.....	7
Related Thermo Scientific Products	7

Introduction

The Thermo Scientific Pierce Magnetic Beads (Table 1) provide a fast and convenient method for both manual and automated magnetic isolation of proteins using affinity binding. Pierce Protein A/G Magnetic Beads are typically used for isolating antibodies from serum, cell culture supernatant or ascites and for immunoprecipitation and co-immunoprecipitation of antigens from cell or tissue extracts. For antibody purification, the beads are incubated with the antibody solution and then magnetically separated from the supernatant. For immunoprecipitation, the beads are added to an antigen-containing sample. The bound antibodies or antigens are dissociated from the beads using an elution buffer. The beads are removed from the solution manually using a magnetic stand or by automation using an instrument such as the Thermo Scientific KingFisher Flex. Automated instruments are especially useful for large-scale screening of multiple samples.

The Pierce Protein A/G Magnetic Beads contain a recombinant Protein A/G (~50,500 Da; apparent molecular weight by SDS-PAGE ~40-45K) that combines the IgG binding domains of both Protein A and Protein G. Protein A/G contains four Fc-binding domains from Protein A and two from Protein G making it a more general and convenient tool for investigating and purifying immunoglobulins. Also, Protein A/G binding to immunoglobulins is not as pH-dependant as Protein A.

Table 1. Characteristics of the Thermo Scientific Pierce Protein A/G Magnetic Beads.

Composition:	Recombinant Protein A/G monolayer covalently coupled to a blocked magnetic bead surface
Magnetization:	Superparamagnetic (no magnetic memory)
Mean Diameter:	1µm (nominal)
Density:	2.0g/cm ³
Bead Concentration:	10mg/mL
Binding Capacity:	55-85µg rabbit IgG/mg of bead

Important Product Information

- Do not centrifuge, dry or freeze the Pierce Magnetic Beads. Centrifuging, drying or freezing will cause the beads to aggregate and lose binding activity.
- To minimize protein degradation, include protease inhibitors (e.g., Thermo Scientific Halt Protease Inhibitor Single-Use Cocktail EDTA-free, Product No. 78425) in preparation of cell lysates.
- A low-pH elution may be used for single-use applications. Optimal time for low-pH elution is 10 minutes; exceeding 10 minutes may result in nonspecific binding and yield reduction.
- When using rabbit antibodies (primary or secondary) in downstream Western blot applications, perform elution in SDS-PAGE sample buffer at room temperature. For all other antibody species, boiling the beads in SDS-PAGE sample buffer is acceptable for single-use applications. Boiling could cause bead aggregation and loss of binding activity.
- Pierce Protein A/G Magnetic Beads are compatible with small-scale antibody purification and immunoprecipitation and analyses by Western blot and mass spectrometry.
- Protein A/G has a broader binding range than either Protein A or Protein G individually. Protein A/G binds to all human IgG subclasses, binds somewhat to IgA, IgE, IgM and, to a lesser extent, IgD. Unlike Protein G, Protein A/G does not bind serum albumin because the gene sequence coding for the albumin-binding site has been eliminated. Protein A/G is effective for mouse monoclonal antibody purification from IgG subclasses because Protein A/G binds all mouse IgG subclasses but does not bind murine IgA, IgM or serum albumin. For more information, see Tech Tip #34: Binding Characteristics for Immunoglobulins and Protein L, A, G and A/G from our website.

Procedure for Manual Antibody Purification

A. Additional Materials Required

- 1.5mL microcentrifuge tubes
- Sample: serum, concentrated cell culture supernatant or concentrated ascites
Note: Samples can be concentrated using the Pierce Concentrators 20mL/20K, Product No. 87751 or 89887A
- Binding/Wash Buffer: Tris-buffered saline (TBS, Product No. 28360) containing 0.05% Tween[®]-20 Detergent
- Elution Buffer: IgG Elution Buffer (Product No. 21004 or 21009) or 0.1M glycine, pH 2-3
- Neutralization Buffer: High-ionic strength alkaline buffer such as a 1M phosphate or 1M Tris; pH 7.5-9
- Magnetic stand (e.g., Thermo Scientific MagnaBind Magnet for 6 × 1.5mL Microcentrifuge Tubes; Product No. 21359)

B. Antibody Purification from Serum, Cell Culture Supernatant or Ascites

Note: To ensure homogeneity, mix the beads thoroughly before use by repeated inversion, gentle vortexing or using a rotating platform.

1. Place 50μL (0.50mg) of Pierce Protein A/G Magnetic beads into a 1.5mL microcentrifuge tube. Add 150μL of Binding/Wash buffer to the beads and gently vortex to mix.
2. Place the tube into a magnetic stand to collect the beads against the side of the tube. Remove and discard the supernatant.
3. Add 1mL of Binding/Wash Buffer to the tube. Invert the tube several times or gently vortex to mix for 1 minute. Collect beads with magnetic stand, then remove and discard the supernatant.
4. Dilute 10μL of sample with 490μL Binding/Wash Buffer.

Note: Sample volume can be modified according to user preference. If the sample volume is < 500μL, dilute it to a final volume of 500μL with Binding/Wash Buffer.

5. Add the diluted sample to the tube containing pre-washed magnetic beads and gently vortex or invert to mix.
6. Incubate the samples at room temperature with mixing for 1 hour.
7. Collect the beads with a magnetic stand, then remove and discard the supernatant.
8. Add 500μL of Binding/Wash Buffer to the tube, mix well, collect the beads with a magnetic stand and discard the supernatant. Repeat this wash twice.

9. Add 100µL of Elution Buffer to the tube, mix well and incubate 10 minutes at room temperature with occasional mixing.
10. Collect the beads with a magnetic stand and then remove and save the supernatant that contains the eluted antibody. To neutralize the low pH, add 10µL of Neutralization Buffer for each 100µL of eluate.

Note: 50µL is the minimum volume of beads recommended for antibody purification.

Procedure for Automated Antibody Purification

A. Additional Materials Required

- KingFisher Flex with 96 deep well head (Product No. 5400630) or KingFisher 96 (Product No. 5400500)
- Microtiter Deep well 96 Plate, V-bottom, polypropylene (100-1000µL; Product No. 95040450)
- KingFisher Flex 96 tip comb for Deep Well Magnets (Product No. 97002534)
- Binding/Wash Buffer: Tris-buffered saline (TBS, Product No. 28360) containing 0.05% Tween-20 Detergent
- Elution Buffer: IgG Elution Buffer (Product No. 21004 or 21009) or 0.1M glycine, pH 2-3
- Neutralization Buffer: High-ionic strength alkaline buffer such as a 1M phosphate or 1M Tris; pH 7.5-9

B. Preparation of Instrument and Plate Set-up

Note: The following protocol is designed for general use with the KingFisher Flex or KingFisher 96 Instrument. The protocol can be modified according to customer needs using the Thermo Scientific BindIt Software provided with the instrument.

1. Download the “Antibody Purification” protocol from the Thermo Fisher Scientific website (www.thermoscientific.com/kingfisher) into the BindIt™ Software on an external computer.
2. Transfer the protocol to the KingFisher Flex or KingFisher 96 from an external computer. See BindIt Software User Manual for detailed instructions on importing protocols.
3. Set up the plates according to Table 2.

Table 2. Pipetting instructions for the antibody purification protocol using the Microtiter Deep Well 96 Plates.

Plate #	Plate Name	Content	Volume
1	Beads	Protein A/G beads	50µL
		Binding/Wash Buffer	150µL
2	Bead Wash	Binding/Wash Buffer	1000µL
3	Bind	Sample	10µL
		Binding/Wash Buffer	490µL
4	Wash 1	Binding/Wash Buffer	500µL
5	Wash 2	Binding/Wash Buffer	500µL
6	Wash 3	Water	500µL
7	Elution	Elution Buffer	100µL
8	Tip Plate	KingFisher Flex 96 tip comb for Deep Well magnets	-

Notes:

- If using less than 96 wells, fill the same wells in each plate. For example, if using wells A1 through A12, use these same wells in all plates.
- To ensure bead homogeneity, mix the vial thoroughly by repeated inversion, gentle vortexing or rotating platform before adding the beads to plate 1.
- Combine the Tip Comb with a Deep Well 96 plate. See KingFisher Flex or KingFisher 96 user manual for detailed instructions.
- Sample volume can be modified according to user preference. If the sample volume is < 500µL dilute it to a final volume of 500µL with Binding/Wash Buffer.

C. Executing the Antibody Purification Protocol on the KingFisher Flex

1. Select the protocol using the arrows on the instrument key pad and press Start. See KingFisher Flex User Manual for detailed information.
2. Slide open the door of the instrument's protective cover.
3. Load the plates into the KingFisher Flex according to the protocol request, placing each plate in the same orientation. Confirm each action by pressing Start.
4. After sample processing, remove plates as instructed by the instrument's display. Press Start after removing each plate.
5. Press Stop after all plates are removed. Upon completion, if desired, neutralize the low pH by adding 10µL of Neutralization Buffer for each 100µL of eluate.

Procedure for Manual Immunoprecipitation**A. Additional Materials Required**

- 1.5mL microcentrifuge tubes
- Binding Buffer: Tris-buffered saline (TBS, Product No. 28360) containing 0.05% Tween-20 Detergent
- Wash Buffer: Tris-buffered saline (TBS, Product No. 28360) containing 0.05% Tween-20 Detergent and 0.5M NaCl
- Low-pH Elution Buffer: IgG Elution Buffer (Product No. 21004 or 21009) or 0.1M glycine, pH 2-3
- Alternative Elution Buffer: SDS-PAGE reducing sample buffer
- Antibody for immunoprecipitation
- Antigen Sample
- Cell Lysis Buffer (used to adjust IP reaction volume)
- Neutralization Buffer: High-ionic strength alkaline buffer such as a 1M phosphate or 1M Tris; pH 7.5-9

B. Immunoprecipitation

Note: This protocol is a general guideline for immunoprecipitation and will require optimization for each application.

1. Combine the antigen sample with 10µg of antibody. Adjust the reaction volume to 500µL with the Cell Lysis Buffer. Incubate the reaction for 1-2 hours at room temperature or overnight at 4°C with mixing.
2. Place 25µL (0.25mg) of Pierce Protein A/G Magnetic Beads into a 1.5mL microcentrifuge tube.
3. Add 175µL of Wash Buffer to the beads and gently vortex to mix.
4. Place the tube into a magnetic stand to collect the beads against the side of the tube. Remove and discard the supernatant.
5. Add 1mL of Wash Buffer to the tube. Invert the tube several times or gently vortex to mix for 1 minute. Collect beads with magnetic stand. Remove and discard the supernatant.
6. Add the antigen sample/antibody mixture to a 1.5mL microcentrifuge tube containing pre-washed magnetic beads and incubate at room temperature for 1 hour with mixing.
7. Collect the beads with a magnetic stand and then remove the flow-through and save for analysis.
8. Add 500µL of Wash Buffer to the tube and gently mix. Collect the beads and discard the supernatant. Repeat wash twice.
9. Add 500µL of purified water to the tube and gently mix. Collect the beads on a magnetic stand and discard the supernatant.
10. Low-pH Elution: Add 100µL of Low-pH Elution Buffer to the tube. Incubate the tube at room temperature with mixing for 10 minutes. Magnetically separate the beads and save the supernatant containing target antigen. To neutralize the low pH, add 10µL of Neutralization Buffer for each 100µL of eluate.

Alternative Elution: Add 100µL of SDS-PAGE reducing sample buffer to the tube and heat the samples at 96-100°C in a heating block for 10 minutes. Magnetically separate the beads and save the supernatant containing target antigen.

Note: If you will be performing a Western blot using rabbit antibodies (primary or secondary) do not heat the samples. Incubate at room temperature for 10 minutes with mixing.

Procedure for Automated Immunoprecipitation

A. Additional Materials Required

- KingFisher Flex with 96 deep well head (Product No. 5400630) or KingFisher 96 (Product No. 5400500)
- Microtiter Deep well 96 Plate, V-bottom, polypropylene (100-1000µL; Product No. 95040450)
- KingFisher Flex 96 tip comb for Deep Well magnets (Product No. 97002534)
- 1.5mL microcentrifuge tubes
- Binding Buffer: Tris-buffered saline (TBS, Product No. 28360) containing 0.05% Tween-20 Detergent
- Wash Buffer: Tris-buffered saline (TBS, Product No. 28360) containing 0.05% Tween-20 Detergent and 0.5M NaCl
- Low-pH Elution Buffer: IgG Elution Buffer (Product No. 21004 or 21009) or 0.1M glycine, pH 2-3
- Alternative Elution Buffer: SDS-PAGE reducing sample buffer
- Antibody for immunoprecipitation
- Antigen Sample
- Cell Lysis Buffer (used to prepare the antigen sample)
- Neutralization Buffer: High-ionic strength alkaline buffer such as a 1M phosphate or 1M Tris; pH 7.5-9)
- Magnetic stand (e.g., Thermo Scientific MagnaBind Magnet for 6 × 1.5mL Microcentrifuge Tubes, Product No. 21359)

B. Instrument Preparation and Plate Set-up

Note: The following protocol is designed for general use with the KingFisher Flex or KingFisher 96 Instrument. The protocol can be modified according to your needs using the BindIt Software provided with the instrument.

1. Combine antigen sample with 2-10µg of immunoprecipitation antibody per sample. Incubate 1-2 hours at room temperature or overnight at 4°C with mixing.
2. Download the “Immunoprecipitation” protocol from the Thermo Fisher Scientific website (www.thermoscientific.com/kingfisher) into the BindIt™ Software on an external computer.
3. Transfer the protocol to the KingFisher Flex or KingFisher 96 from an external computer. See BindIt Software User Manual for detailed instructions on importing protocols.
4. Set up plates according to Table 3.

Table 3. Pipetting instructions for the immunoprecipitation protocol using the Microtiter Deep Well 96 Plates.

Plate #	Plate Name	Content	Volume	Time/Speed
1	Beads	Protein A/G Beads	25µL	5 seconds
		Binding Buffer	175µL	
2	Bead Wash	Binding Buffer	1000µL	1 minute/Slow
3	Bind	Antibody/Antigen Sample	500µL	1 hour/Slow
4	Wash 1	Wash Buffer	500µL	30 seconds/Slow
5	Wash 2	Wash Buffer	500µL	30 seconds/Slow
6	Wash 3	Ultrapure Water	500µL	30 seconds/Slow
7	Elution	Elution Buffer	100µL	10 minutes/ Medium
8	Tip Plate	KingFisher Flex 96 Tip Comb for Deep Well Magnets	-	10 seconds/Fast

Notes:

- If less than 96 wells are used, fill the same wells in each plate. For example, if using wells A1 through A12, use these same wells in all plates.
- To ensure bead homogeneity, mix the vial thoroughly by repeated inversion, gentle vortexing, or rotating platform before adding the beads to Plate 1.

- Combine the Tip Comb with a Deep Well 96 Plate. See the instrument user manual for detailed instructions.
- The beads can be eluted into 100µL of 0.1M glycine, pH 2-3 or 100µL SDS-PAGE reducing sample buffer. If using SDS-PAGE reducing sample buffer in a heated elution, install the KingFisher Flex or 96 Heating Block (see manual for proper installation) to heat samples at 96-100°C for 10 minutes.
- If you select SDS-PAGE reducing sample buffer for elution and will be performing a Western blot using rabbit antibodies (primary or secondary) do not heat the samples. Incubate at room temperature for 10 minutes.
- If low-pH elution buffer is selected for elution, neutralize the pH using 10µL Neutralization Buffer for each 100µL of eluate upon run completion.
- To limit evaporation, select “Mix” and “Slow” speed under the subheading “Heating Action”

C. Executing Automated Immunoprecipitation Protocol

1. Select the protocol using the arrow keys on the instrument keypad and press Start. See the KingFisher Flex or KingFisher 96 User Manual for detailed information.
2. Slide open the door of the instrument’s protective cover.
3. Load plates into the instrument according to the protocol request, placing each plate in the same orientation. Confirm each action by pressing Start.
4. After the samples are processed, remove the plates as instructed by the instrument’s display. Press Start after removing each plate. Press Stop after all the plates are removed.

Troubleshooting

Problem	Possible Cause	Solution
Low amount of protein was recovered	The protein degraded	Add protease inhibitors
	Not enough magnetic beads were used	Increase the amount of magnetic bead used for capture
	Sample had an insufficient amount of target protein	Increase amount of antigen sample
Protein does not elute	Elution conditions were too mild	Increase incubation time with elution buffer or use more stringent elution buffer
Bands at ~50kDa appeared on the Western blot	Elution was performed in Lane Marker Sample Buffer at temperatures above room temperature and rabbit antibody was used in the Western blot detection	Perform elution at room temperature when using a rabbit antibody for Western blot detection
Multiple nonspecific bands	Nonspecific protein bound to the magnetic beads	Add 50-350mM of NaCl to the Binding/Wash and Elution Buffers
Recovered protein was inactive	Elution conditions were too stringent	Use a milder elution buffer (e.g., Thermo Scientific Gentle Elution Buffer, Product No. 21034)
Magnetic beads aggregated	Magnetic beads were frozen or centrifuged	Handle the beads as directed in the instructions
	Buffer was incompatible with magnetic beads	

Additional Information

Visit the www.thermoscientific.com/pierce for additional information relating to this product including the following:

- Frequently Asked Questions
- Tech Tip #43: Protein Stability and Storage
- Tech Tip #34: Binding Characteristics for Immunoglobulins and Protein L, A, G and A/G
- Visit www.thermoscientific.com/kingfisher for information on the KingFisher Products
- In the U.S.A., purchase KingFisher Supplies from VWR. Contact your local Thermo Fisher Scientific office to purchase KingFisher Supplies outside the U.S.A.

Frequently Asked Questions for the KingFisher Instrument

Question	Answer
Which plates are compatible with KingFisher Flex and KingFisher 96 instruments?	The KingFisher Flex instrument is compatible with the KingFisher 24 Deep Well plates, Microtiter Deep Well 96 plates, KingFisher 96 and 96 PCR plates.
Is it possible to concentrate samples during the run?	Both deep-well plates and KingFisher 96 plates can be used during the same run. Therefore, it is possible to start the processing using larger volumes (in a deep well plate) and elute the purified sample to a smaller volume (in a KingFisher 96 plate)
Is it possible to heat the samples during the run?	The heating block is located inside the instrument and can be used automatically during the sample process. All plates compatible with the KingFisher Flex instrument can be heated using specially designed, interchangeable heating blocks.
Why do the beads stick to the plastic tips and wells or the eluted protein sticks to the wells?	Proteins conjugate to beads and eluted proteins can nonspecifically bind to plastics. Adding detergent to Binding/Wash Buffer prevents the protein conjugated to the bead from sticking (0.05%-0.1% Tween-20 Detergent). Also include a small amount of detergent in the elution buffer (e.g., 0.05% Tween-20 Detergent) or silanize the elution plate.
Are the reagent volumes in each well critical?	For best results, keep the specified volumes within defined limits to avoid spillover.

Related Thermo Scientific Products

88804	Pierce Protein A/G Magnetic IP/Co-IP Kit
88816, 88817	Pierce Streptavidin Magnetic Beads
88821, 88822	Pierce Glutathione Magnetic Beads
24615	Imperial Protein Stain, 1L, sufficient for up to 50 mini gels
34075	SuperSignal® West Dura Extended Duration Substrate, 100mL
25200-25244	Precise Gels, see catalog or website for a complete listing
88811, 88812	Pierce Magnetic TiO₂ Phosphopeptide Enrichment Kit

Tween® is a registered trademark of ICI Americas.

This product ("Product") is warranted to operate or perform substantially in conformance with published Product specifications in effect at the time of sale, as set forth in the Product documentation, specifications and/or accompanying package inserts ("Documentation") and to be free from defects in material and workmanship. Unless otherwise expressly authorized in writing, Products are supplied for research use only. No claim of suitability for use in applications regulated by FDA is made. The warranty provided herein is valid only when used by properly trained individuals. Unless otherwise stated in the Documentation, this warranty is limited to one year from date of shipment when the Product is subjected to normal, proper and intended usage. This warranty does not extend to anyone other than the original purchaser of the Product ("Buyer").

No other warranties, express or implied, are granted, including without limitation, implied warranties of merchantability, fitness for any particular purpose, or non infringement. Buyer's exclusive remedy for non-conforming Products during the warranty period is limited to replacement of or refund for the non-conforming Product(s).

There is no obligation to replace Products as the result of (i) accident, disaster or event of force majeure, (ii) misuse, fault or negligence of or by Buyer, (iii) use of the Products in a manner for which they were not designed, or (iv) improper storage and handling of the Products.

Current product instructions are available at www.thermoscientific.com/pierce. For a faxed copy, call 800-874-3723 or contact your local distributor.

© 2011 Thermo Fisher Scientific Inc. All rights reserved. Unless otherwise indicated, all trademarks are property of Thermo Fisher Scientific Inc. and its subsidiaries. Printed in the USA.